



Bid Number/बोली क्रमांक (बिड संख्या):  
GEM/2025/B/6161455  
Dated/दिनांक : 21-05-2025

### Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	16-06-2025 15:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	16-06-2025 15:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Labour And Employment
Department Name/विभाग का नाम	Na
Organisation Name/संगठन का नाम	Employees State Insurance Corporation (esic)
Office Name/कार्यालय का नाम	Esic Medical College & Hospital, Gulbarga
क्रेता ईमेल/Buyer Email	msomch-gb.kar@esic.gov.in
Total Quantity/कुल मात्रा	2
Item Category/मद केटेगरी	Cardiac TMT Machine (V2) (Q2)
Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का)	6 Lakh (s)
OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)	24 Lakh (s)
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	3 Year (s)
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes

<b>Bid Details/बिड विवरण</b>	
<b>Document required from seller/विक्रेता से मांगे गए दस्तावेज़</b>	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
<b>Do you want to show documents uploaded by bidders to all bidders participated in bid?/</b>	No
<b>Past Performance/विगत प्रदर्शन</b>	10 %
<b>Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया</b>	Yes
<b>RA Qualification Rule</b>	50% Lowest Priced Technically Qualified Bidders
<b>Comprehensive Maintenance Charges Required</b>	Yes
<b>Type of Bid/बिड का प्रकार</b>	Two Packet Bid
<b>Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय</b>	2 Days
<b>Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b>	No
<b>Evaluation Method/मूल्यांकन पद्धति</b>	Total value wise evaluation
<b>Arbitration Clause</b>	No
<b>Mediation Clause</b>	No

#### **EMD Detail/ईएमडी विवरण**

Advisory Bank/एडवाइजरी बैंक	State Bank of India
EMD Amount/ईएमडी राशि	24000

#### **ePBG Detail/ईपीबीजी विवरण**

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	3.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	62

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the

relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

**Beneficiary/लाभार्थी :**

Dean

Esic Medical College & Hospital, Gulbarga, NA, Employees State Insurance Corporation (ESIC), Ministry of Labour and Employment

(Esi Fund Account No 1)

**MII Purchase Preference/एमआईआई खरीद वरीयता**

MII Purchase Preference/एमआईआई खरीद वरीयता	Yes
--	-----

**MSE Purchase Preference/एमएसई खरीद वरीयता**

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
5. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
6. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its

subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

7. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

8. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

9. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

10. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available. Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

## **Cardiac TMT Machine (V2) ( 2 pieces )**

**(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)**

### **Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL FEATURES	Product Description	Cardiac TMT Machine with Monitor and Printer
	Product components	Heavy-duty medical treadmill with online UPS for the whole system, Computer system, printer, Voltage stabilizer and software along with trolley system to mount computer & printer
	Facility, whereby data should be able to be acquired and digitized near the patient, providing better ECG signal	Yes
TREADMILL SPECIFICATIONS	Treadmill Type	Four tier heavy duty noise less medical treadmill with casters having locking arrangement
	Compatible online UPS for the whole system with minimum 3 hrs back up time	Yes
	Stop/start button on treadmill for emergency stop	Yes
	Emergency stop button (ESB) must have the ability to be located in the location of choice by the end user	Yes
	Treadmill should have zero start	Yes
	Automatic operation from the main unit as per test protocol	Yes
	Fold down hand rails on treadmill for allowing easy and convenient assemble and installation	Yes
	Length of Treadmill Floor (mm)	1900 to 2000mm
	Width of Treadmill Floor (mm)	700 to 800mm
	Length of walking area on treadmill (mm)	1500 to 1600mm
	Width of walking area on treadmill (mm)	500 to 600mm
	Maximum Patient Load Capacity of Treadmill (Kg)	150 Kg or more, 200 Kg or more Or higher
	Speed range of Treadmill	0.1 to 14 km/Hr or higher range
	Inclination (continuously variable)	0 to 25 or more
	Treadmill motor	AC Motor
	Power requirements	220 - 240 VAC @ 50-60 Hz

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Double sided polish on walking surface of treadmill for prolonged product life span	Yes
	Remote keyboard operation to start, stop and increase or decrease elevation and speed of treadmill	Yes
	Self-aligning anti skid conveyor belt	Yes
	Communication/Connectivity	RS232, USB
	Enhanced tolerance to electrical changes for greater patient safety	Yes
	Self-calibrating system for speed	Yes
	Inbuilt digital control issues, precise speed and grade control in the system	Yes
	Delivers high torque even at low speeds which ensures no lag when the foot strikes the belt	Yes
	Provide smooth power and accuracy even at slow speeds	Yes
	Normal, slow deceleration and emergency stop modes on treadmill	Yes
	Compatible Voltage Stabilizer provided (Minimum 3KVA)	Yes
COMPUTER SYSTEM SPECIFICATIONS	Type of Monitor	LCD, LED Or higher
	<b>Touch Screen Monitor</b>	No
	Size of Monitor (inch)	24 Or higher ( <b>inch</b> )
	RAM size	8 GB Or higher
	Hard disk space	1 TB Or higher
	Processor	Intel core i5
	Operating System	Windows 10 or Higher
	CD ROM Drive	Yes
	Number of serial port	2
	Number of USB Ports	4
	Alphanumeric multimedia PC keyboard with systems dedicated keypad	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Standard PC mouse	Yes
	Built in uninterrupted power supply for minimum of 3 minutes	Yes
	QRS triggered output	Yes
	Parameters Displayed by the system	12 Lead raw ECG, 12 lead ST level & Slops, Enlarged QRS complex with maximum ST level, Exercise time, Target and Max Heart Rate with % of target achievement, Heart Rate & METS trends, NIBP Trends, Speed of treadmill, Protocol Name, Protocol stage, Patient info
	Built in thermal printer with compatible additional laser black and white printer to print report on A4 Size thermal and A4 size ordinary bond papers respectively	Yes
	Minimum Printing Speed of laser printer	20 pages / minute
	Minimum Paper tray capacity of laser printer	250 Sheets
	Print Resolution	1200 x 600 dpi
	Duplex Printing	No
	Connectivity / interface of Laser printer	USB
	Supported paper sizes	A4
TRANSMISSION OF DATA	Transmission of data available through	USB, PDF
	Advanced signal processing algorithms to provide exceptional accuracy in beat detection and noise rejection resulting in superior identification of events	Yes
	Unique source consistency filter, reducing noise and baseline artifact while preserving a diagnostic quality ECG signal	Yes
	Provision to eliminate artifacts due to respiration, muscle rub and AC interference, baseline wandering without compromising and distortion in ST segment changes	Yes
PROTOCOLS	Test Protocols include Bruce, Modified Bruce, Balke, Ellestad, Naughton protocols	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Facility to add user defined protocols	Yes
SYSTEM SOFTWARE FEATURES	Licensed copy of windows operating system	Yes
	Provision to record resting ECG & Exercise ECG	Yes
	Resting ECG software with detailed automatic interpretation statement for all 12 or 16 lead resting ECG should be present	Yes
	ECG data management software for storage and review	Yes
	Facility to review, edit and add ECG from full discloser storage post exam	Yes
	50mm sweep speed selection for ECG display in the system	Yes
	Enlarged QRS complexes with superimposition technique	Yes
	Automatic BP Measurement System integrated with the stress test system	Yes
	Provision of real time average complexes	Yes
	Provision of real time ST Analysis in bar graphs	Yes
	Provision of Re analysis of final summary report	Yes
	System should be able to present ST levels for 12 average complexes, slope for 12 average complexes and ST Profile	Yes
	Provision of online ST measurement adjustment	Yes
	Provision of reanalyzing the possibility of complete exercise test with ST/HR and NIBP trend graphically	Yes
	SCD predictive tools like T-wave alternans, ST /HR hysteresis analysis	Yes
The system should have a Borg scale, symptom and point criteria table	Yes	

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Full disclosure review/playback/scroll back and addition of past ECG events	Yes
	ST level and slope data for lead and worst case average beat update should be available continuously during the test	Yes
	Beat by beat review mode to allow full disclosure review during and post exam	Yes
	QRS signal averaging software, QT dispersion software, XYZ vector-cardiography software provided as standard part of the unit	Yes
	Facility to do reanalysis of scored ECG by changing line measurement point title ST and J	Yes
	The system should present comprehensive final report on minute by minute record of ST segment trend	Yes
	Facility of automatic arrhythmia detection	Yes
	Auto calibration with digital micro process control in the system and should also be manually operable	Yes
	Facility of pacemaker rhythm detection	Yes
	System should be able to view comprehensive ST segment and morphology analysis in an anatomically intelligent format	Yes
	ST profile, ST elevation, ST depression, heart rate, Mets, NIBP, ST index should be able to be clearly analyzed	Yes
	The unit should be able to quickly identify anomalies with the dynamic zoom ST display	Yes
	The system should be able to view ST maps, patented visual anatomical representations of ST deviations in-frontal planes	Yes
	Monitoring of ST changes with auto comparison of current and reference beats	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	The system should be able to display the traces incrementally	Yes
	Provision of updating ECG and ST segment changes and compare them on a pre-exercise and during exercise basis on a high resolution display	Yes
	Compact digital acquisition module that virtually eliminates non-cardiac electrical noises	Yes
	Provision of alarm on HR, BP & ST level	Yes
	Facility to mark the ECG strip to enter comments at any stage	Yes
DATA ACQUISITION AND PROCESSING	Data Acquisition	12 Leads ECG Simultaneously and Display 3x4, 4x6, 6x12, 12 lead data indicate significant changes
	A/D	12 bits or more
	Sampling Rate	300/Sec/channel or more
	Input impedance should be more than 90M Ohms	Yes
	CMRR Should be more than 100 dB	Yes
	Frequency response	0.05Hz to 100Hz
	Patient leakage Should be less than 100?A	Yes
	The system should have baseline correction	Yes
	Digital filters	50Hz, 20Hz, 35Hz
REPORTING	Facility to get system generated auto report	Yes
	Analysis should work on total of atleast 5 protocols	Yes
	No startup delay in the system ie instant ECG Reports at the push of a button should be available	Yes
	Final report with configurable lead groups	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Patient report include protocol, Indications, medications, target heart rate reasons for end, symptoms, diagnosis, notes and conclusions with space for personnel and physician electronic signature	Yes
	Examination summary include heart rate/BP/Workload trends page, ST level trends/ST slope trends, Average QRS by stage or by minute, Mets, Maximum predicted heart rate, THR formula selection	Yes
	Ability to select final report segments including patient information, exam summary, rate, BP, work trends, ST level trends, ST slope trends, average QRS and events should be provided in the system	Yes
	Events should include 12 lead ECG analysis for arrhythmias and user added ECGs	Yes
	Provision of customizable system generated summary report to determine the contents of report with data storage facility	Yes
	Provision of Password protection	Yes
	Option to create network review workstations for viewing, editing and report printing of the stress test examinations should be there	Yes
ACCESSORIES	Standard Accessories, consumables provided	2 sets of 12 lead ECG patient cables, 100 Electrodes, 10 rim A4 Reporting papers
	Detachable/replaceable cables and leads	Yes
	Minimum warranty(months) on accessories i e lead set	12 Or higher <b>(year)</b>
	Appropriate sized cuffs of small, medium and large should be provided with the system	Yes
	Trolley system to mount computer & printer	Yes
STORAGE OF RECORDS	Storage of atleast 100 patients records	Yes

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Full disclosure archive to external drive	Yes
	Events/12 lead ECGs/Full disclosure	Yes
CERTIFICATION & REPORTS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid drug license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Manufacturing unit certification	ISO:13485 (Latest)
	Availability of Test Report for product as per Medical Device Rules (MDR) 2017 as amended till date	Yes
	Compliance to Safety Standards	IEC-60601-2-25 or equivalent BIS Standard
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement	Yes
WARRANTY	<b>Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)</b>	3, 4, 5 Or higher (year)
MISCELLANEOUS PARAMETERS	OEM/Reseller shall ensure uninterrupted availability of all spares for 10 years	Yes
	Availability of toll free facility for technical support maintained by OEM or authorized agencies	Yes
	User/Technical/Maintenance manuals to be supplied in English in hard and soft copy	Yes
	Details of equipments and procedures required for local calibration and routine maintenance to be supplied and advanced maintenance task documentation also to be furnished	Yes

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	List of important spares and accessories, with their part numbers to be supplied to the buyer at the time of supplying the equipment	Yes
	Installation and Demonstration of equipment and training to be provided after completing supplies before acceptance	Yes
	Principal Manufacturer must have direct Presence/approved service center In India	Yes

Comprehensive Maintenance	
Warranty of required product	5 Year
Comprehensive Maintenance Duration (Post Warranty)	5 Year

\*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

#### Additional Specification Parameters - Cardiac TMT Machine (V2) ( 2 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
1. Communication/Connectivity 2. Type of Monitor	1. To be read as USB instead of RS232, USB 2. To be read as LCD instead of LCD, LED Or higher
3. WARRANTY	3.Needs to be read as 5 Years and above

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Vini Kaila	585106,ESIC MEDICAL COLLEGE & HOSPITAL SEDAM ROAD KALABURAGI-585106	2	15

**Special terms and conditions-Version:1 effective from 06-10-2023 for category Cardiac TMT Machine (V2)**

1.
  1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
  2. The sellers are registered on GeM based on the submission of valid Drug License and self declaration of product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
  3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
  4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
  5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.
  6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
  7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
  8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
  9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
  10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
  11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation &

commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.

12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

## **Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**

### **1. Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

### **2. Generic**

Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.

### **3. Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

### **4. Service & Support**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

### **5. Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

### **6. Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

### **7. Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

## 8. **Certificates**

ISO 9001: The bidder or the OEM of the offered products must have ISO 9001 certification.

## 9. **Certificates**

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

## 10. **Warranty**

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

## 11. **Warranty**

Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The comprehensive warranty shall be covering the following scope

Spares, Labour, repairs and rectification or replacement of the components or up gradation of software or calibration

(Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)

## 12. **Warranty**

Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

## 13. **Forms of EMD and PBG**

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

ESIC Fund Account No 1  
payable at  
Kalaburagi

Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

## 14. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

1. The tenderer must enclose an undertaking in the letter head as mentioned in Annexure - I (ATC Attachment)
2. The tenderer must enclose an undertaking in the letter head as mentioned in Annexure - II (ATC Attachment)

3. It is mandatory for all the bidders to provide a working physical demonstration before the Technical Evaluation Committee members nominated by the Competent Authority at ESIC Medical College & Hospital, Kala buragi for qualifying in the bid or else will be liable for rejection. The cost of the demonstration should be borne by the bidder. The recommendation of the Technical Evaluation Committee will be binding to qualify technically.
4. Standby equipment needs to be provided if the equipment cannot be repaired and restored to its functionality within 48-72 hours of intimation of repair/ maintenance during warranty or comprehensive maintenance contract period.
5. The equipment up gradation of software and or calibration needs to be covered free of cost by the firm/vendor during the warranty and Comprehensive Maintenance Contract period as far as possible within the premises of hospital/ institution.
6. Consumable accessories, if any, not covered in warranty/CMC should be clearly specified.
7. Prices of consumables and accessories should be quoted separately and frozen for the period of warranty and CMC
8. It is mandatory to provide the Satisfactory certificate from Central/State Govt. organization/PSU/Public Listed Company
9. Bidders may please note that GeM is capturing and showing the IP addresses used by the Buyer and the Bidder(s) /Seller(s). The received bids having matching/common IP address with Bidder/Seller(s) or Buyer shall be out rightly rejected & shall not be considered for further evaluation.
10. Bidder should submit ISO/BIS/CDSCO/Any Indian Certifications

15. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

16. **Generic**

**Bidder financial standing:** The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

17. **OEM**

**IMPORTED PRODUCTS:** In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

18. **Warranty**

Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

19. **Generic**

Bidders shall quote only those products (Part of Service delivery) in the bid which are not obsolete in the market and has at least 7 years residual market life i.e. the offered product shall not be declared end-of-life by the OEM before this period.

20. **Generic**

**Manufacturer Authorization:** Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

21. **Certificates**

To be eligible for award of contract, Bidder / OEM must possess following Certificates / Test Reports on the date of bid opening (to be uploaded with bid):

## **Disclaimer/अस्वीकरण**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

## **Additional Clause For Comprehensive Maintenance Charges**

1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98%

uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.

2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.

3.GST shall be included in the CMC Charges quoted.

4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.

5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.

6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No” options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available.Under this option up to 10 years can be chosen for CMC charges beyond warranty period.

7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.

8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period- Percentage to be indicated- A1

CMC charges for 2nd year after warranty period- Percentage to be indicated- A2

CMC charges for 3rd year after warranty period - Percentage to be indicated- A3

CMC charges for 4th year after warranty period - Percentage to be indicated- A4

CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.

(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C + C * \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$  and so on

C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1,A2, A3, A4& A5 shall depend on how many years CMC selected. For 3 year CMC, only A1,A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4.CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5.The CMC charges shall be offered within range of 3 to 10% of cost of equipment.

9.Since CMC charges are to be paid only later for each year during CMC period,applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10.Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges.The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification.Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11.In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis.The equipment cost and CMC charges (year to year) shall be matched individually.

12.The CMC Contract shall be an offline contract to be handled by buyer.The payment of CMC will be made on

quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13.CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

**All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.**

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---